

Anti-Rabies Virus (RV) Glycoprotein (G) Human IgG ELISA Kit

Overview

Catalog No.	AK584018
Specifications	Glycoprotein
Applications	ELISA
Stability and Storage	The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 10% prior to the expiration date under appropriate storage condition.
Detection method	Colorimetric
Sample type	Plasma, Serum
Assay type	Quantitative
Sensitivity	1.09 ng/mL
Range	1.56 - 100 ng/mL
Recovery	80-120%
Shipping	2-8 °C
Note	For Research Use Only.