

## SARS-CoV-2 Surrogate Virus Neutralization Test Kit (Omicron)

## **Overview**

Catalog No. VK726028

**Specifications** SARS-CoV-2 Neutralization Antibody

**Applications** ELISA

The stability of ELISA kit is determined by the loss rate of activity. The loss

**Stability and Storage** rate of this kit is less than 10% prior to the expiration date under

appropriate storage condition.

**Detection method** Colorimetric

Sample type Plasma, Serum

**Assay type** Quantitative

**Sensitivity** 78.1 ng/mL

**Range**  $0.156 - 10 \,\mu \text{g/mL}$ 

**Recovery** 80-120%

Shipping 2-8 °C

**Note** For Research Use Only.